

	CSIR - NATIONAL GEOPHYSICAL RESEARCH INSTITUTE
	(Council of Scientific & Industrial Research)
	UPPAL ROAD, HYDERABAD-500 007, Telangana, INDIA
	Tel: 91 40 27012370/27012486; Fax:+91 40 27171564; Email: purchase@ngri.res.in

### Technical clarifications

against tender No. 57(192)2015-PUR/T-142 dated 09/03/2016  
for Supply, Installation and Commissioning of Broad Band Seismograph

S. No	NGRI Specifications	Clarification sought	NGRI Reply
	Tender document – Part-B – Data Acquisition System		
11	Status display: LCD display for power, data acquisition and state of health should be on the top / front panel for the DAS	As this LCD display on the top of the recorder is manufactured by limited people in the world.	<p>LCD display provides the alphanumeric display of the absolute values of the battery voltage, data acquisition status and state of health parameters which are more informative than LED bulb display.</p> <p>LCD display is useful for the operator / field staff to know the state of health information locally without connecting any external device such as palmtop or laptop computer.</p> <p>Accordingly LCD Display is acceptable.</p>

Controller of Stores & Purchase